

# codex alimentarius commission

# E



FOOD AND AGRICULTURE  
ORGANIZATION  
OF THE UNITED NATIONS

WORLD  
HEALTH  
ORGANIZATION



JOINT OFFICE: Viale delle Terme di Caracalla 00153 ROME Tel: 39 06 57051 www.codexalimentarius.net Email: codex@fao.org Facsimile: 39 06 5705 4593

**Agenda Item 3**

**CX/CF 07/1/2  
March 2007**

**JOINT FAO/WHO FOOD STANDARDS PROGRAMME  
CODEX COMMITTEE ON CONTAMINANTS IN FOODS  
First Session  
Beijing, China, 16 - 20 April 2007**

**MATTERS REFERRED TO THE COMMITTEE BY THE CODEX ALIMENTARIUS  
COMMISSION AND/OR OTHER CODEX COMMITTEES / TASK FORCES**

**A. MATTERS ARISING FROM THE 29<sup>th</sup> SESSION OF THE CODEX ALIMENTARIUS  
COMMISSION (Geneva, Switzerland, 2006)**

**Matters for action**

**Terms of Reference of the Codex Committee on Contaminants in Foods**

1. The Commission adopted the terms of reference for the Codex Committee on Contaminants in Foods. The Commission agreed that the Committee should review its terms of reference at its first session<sup>1</sup>.
2. The Committee is therefore invited to review its terms of reference (reproduced below) and to inform the Commission as to whether it is satisfied with them, together with any observations as appropriate.
  - (a) to establish or endorse permitted maximum levels or guidelines levels for contaminants and naturally occurring toxicants in food and feed;
  - (b) to prepare priority lists of contaminants and naturally occurring toxicants for risk assessment by the Joint FAO/WHO Expert Committee on Food Additives;
  - (c) to consider methods of analysis and sampling for the determination of contaminants and naturally occurring toxicants in food and feed;
  - (d) to consider and elaborate standards or codes of practice for related subjects; and
  - (e) to consider other matters assigned to it by the Commission in relation to contaminants and naturally occurring toxicants in food and feed.

<sup>1</sup> ALINORM 06/29/41, para. 26 and Appendix III

**Matters for information****Approval of new work for the elaboration of new standards and related texts**

3. The Commission approved the following items as new work of the Committee<sup>2</sup>:
- Revision of the Preamble of the Codex General Standard for Contaminants and Toxins in Food (Job Code: N04-2006);
  - Proposed Draft Code of Practice for the Prevention and Reduction of Ochratoxin A Contamination in Wine (Job Code: N05-2006);
  - Proposed Draft Code of Practice for the Reduction of Acrylamide in Food (Job Code: N06-2006); and
  - Proposed Draft Code of Practice for the Reduction of Contamination of Food with Polycyclic Aromatic Hydrocarbons (PAH) from Smoking and Direct Drying Processes (Job Code: N07-2006).

**Adoption of the Draft Standards and Related Texts at Step 8 or Step 5/8**

4. The Commission adopted the following maximum levels and other texts at Step 8 or Step 5/8<sup>3</sup>:
- Maximum Level for Lead in Fish;
  - Maximum Levels for Cadmium in marine bivalve molluscs (excluding oysters and scallops), in cephalopods (without viscera) and in polished rice;
  - Appendix to the Codex Code of Practice for the Prevention and Reduction of Aflatoxins Contamination in Tree Nuts – Additional Measures for the Prevention and Reduction of Aflatoxins Contamination in Brazil nuts;
  - Code of Practice for the Prevention and Reduction of Dioxin and Dioxin-like PCB Contamination in Food and Feeds;
  - Revision of the Guideline Levels for Radionuclides in Foods Contaminated Following a Nuclear or Radiological Emergency for Use in International Trade; and
  - Codex General Standard for Contaminants and Toxins in Foods, including Schedule I.

**Adoption of the Draft Standards and Related Texts at Step 5**

5. The Commission adopted below the draft maximum levels at Step 5 and advanced them to Step 6<sup>4</sup>:
- Draft Maximum Level for Total Aflatoxins Almonds, Hazelnuts and Pistachios “ready to eat”; and
  - Draft Maximum Levels for Tin in Canned Foods (other than beverages) and in Canned Beverages.

**Revocation of the standards and related texts**

6. The Commission agreed to the revocation of the following text<sup>5</sup>:
- List of Individual Codex Maximum Levels and Guideline Levels for Contaminants and Toxins.

**FAO and WHO Expert Consultation on the health risks associated with methylmercury and dioxins and dioxin – like PCBs in fish and the health benefits of fish consumption**

7. The Commission agreed to request FAO/WHO to consider convening an expert consultation on the health risks associated with the consumption of fish and other seafood and the health benefits of fish and other seafood consumption, with the detailed terms of reference as proposed by the CCFAC<sup>6</sup>.

<sup>2</sup> ALINORM 06/29/41 paras 124-127 and Appendix VIII

<sup>3</sup> ALINORM 06/29/41 para.36 and Appendix IV

<sup>4</sup> ALINORM 06/29/41 para. 97 and Appendix V

<sup>5</sup> ALINORM 06/29/41 para. 120 and Appendix VII

<sup>6</sup> ALINORM 06/29/41 para. 195, ALINORM 06/29/12 para. 191

**B. MATTERS ARISING FROM OTHER CODEX COMMITTEES AND TASK FORCES****Matters for action****Methods of analysis for the determination of dioxins and PCBs<sup>7</sup>**

8. The Codex Committee on Methods of Analysis and Sampling (CCMAS), at its 27th Session (May 2006) agreed to request the Codex Committee on Contaminants in Foods (CCCF) to provide precise information on the range of levels to be considered as well as the matrices for which these levels are to be applied. And the CCMAS agreed to forward the document on Methods of Analysis for the Determination of Dioxins and PCBs for their information and that new work in this regard would only be resumed pending a reply from the CCCF.

9. The Committee is therefore invited to consider the document on the methods of analysis for the determination of dioxins and PCBs<sup>8</sup>, which is reproduced in Annex to this document, in response to the above question.

**Matters for information**

10. No information has been forwarded from other codex committees and task forces.

---

<sup>7</sup> ALINORM 06/29/23 paras 10-11 and 92-97

<sup>8</sup> CX/MAS 06/27/8 (in English only)

## METHODS OF ANALYSIS FOR THE DETERMINATION OF DIOXINS AND PCBS

### Background

At the 26<sup>th</sup> session of the Codex Committee on Methods of Analysis and Sampling (CCMAS) in Budapest, Hungary, 4 – 8 April 2005, the Committee decided to inform the Codex Committee on Food Additives and Contaminants (CCFAC) about the status of its work on methods of analysis for dioxins (CX/FAC 06/38/2-Add.1).

The Committee requested the delegation of Germany to revise the paper with the view of converting the already reported methods used for the determination of dioxins and related compounds into criteria. Furthermore all governments and international organisations were again invited to provide information on currently used methods for dioxin analysis to the delegation of Germany before next session (ALINORM 05/28/23 para 123).

This initiative has its origin in a request of the Codex Committee on Food Additives and Contaminants (CCFAC) concerning information on methods of analysis for dioxins. CCFAC is currently drafting a Code of Practice for the Prevention and Reduction of Dioxin and Dioxin-like PCB Contamination in Foods and Feed which was considered (at Step 3) at the 38th session in The Hague, the Netherlands, 24-28 April 2006.

Some delegations in CCFAC expressed the view that although there were no limits in Codex for dioxins, it would be useful to consider the selection of appropriate methods in the Committee, taking into account the work underway in different international organisations.

A lot of work has been done by the Joint FAO/WHO Expert Committee on Food Additives (JECFA) ([www.who.int](http://www.who.int)).

Due to the heterogeneity in analytical approach JECFA and the European Community do not consider a standardisation of those methods.

Up-to-date there exists no official internationally agreed method for the determination of dioxins and related compound.

### Methods used to determine dioxins and related compounds

PCDDs/PCDFs are normally found as complex mixtures in varying composition in different matrices. Their identification and quantification requires a highly sophisticated analysis, because it is necessary to separate the toxic (17 congeners with 2,3,7,8- chlorine substitution) from the less-toxic congeners. Usually, PCDDs/PCDFs are determined by capillary-GC/MS (gas chromatography / mass spectrometry) methods.

In the past, PCB analyses mainly focused on the determination of total PCBs or marker congeners (PCBs 28, 52, 101, 138, 153 and 180, which are the predominant PCB congeners found in humans and food stuffs of animal origin). However, the toxicity of these PCB congeners appears to be relatively low. Based on the available toxicological information, the non-ortho PCBs 77, 81, 126 and 169 and the mono-ortho congeners 105, 114, 118, 123, 156, 157, 167 and 189 were assigned a toxic equivalency factor (TEF) by a WHO expert group in 1997 and have to be analysed to determine the PCB-TEQ content. Data for these dioxin-like PCB congeners are still scarce. Due to their chemical and physical properties mono-ortho PCBs and non-ortho PCBs have to be determined separately in most cases. Reliable determinations of non-ortho PCBs in food have been performed by high-resolution MS, as collaborative studies demonstrate.

## GC-HRMS

Gas Chromatography combined with High Resolution Mass Spectrometry is currently the only technique able to provide the required sensitivity and selectivity for analysis and detection of dioxins and dioxin-like PCBs. Contrary to the biological screening techniques (that measure the sum of the toxic dioxins in the sample), GC-HRMS allows to separate and detect the individual dioxins that contribute to the sum of toxic dioxins in a sample. The main difference of HRMS compared with low resolution MS is the fact that HRMS has significant more separating power (resolution) to allow separation of the dioxin-borne ions from other interfering ions. In that way HRMS is able to detect dioxins at very low levels without interference from other compounds. To assure reliable detection, generally, quantification is performed by addition of isotope-labelled  $^{13}\text{C}_{12}$  analogues of the individual dioxins which are added to the sample before analysis and detected separately by the HRMS.

### *GCxGC*

In environmental analysis complex mixtures like dioxins, PCBs and brominated flame retardants require high separating power to enable the detection of all individual compounds. Conventional single column capillary gas chromatography offers much separation but often suffers from co-eluting compounds or (unknown) interferences.

In comprehensive two-dimensional gas chromatography (GCxGC) two independent separations are applied to an entire sample. The sample is first separated on a normal-bore capillary column under programmed-temperature conditions. The effluent of this column then enters a thermal (or cryo) modulator, which traps each subsequent small portion of eluate, focuses these portions and releases the compounds into a second column for further separation. The second separation is made to be fast enough (e.g. 5 – 10 s) to permit the continual introduction of subsequent, equally small fractions from the first column without mutual interference.

## Methods reported by Member Countries

Only two countries have provided further information on methods for the detection and identification of dioxins and related compounds which have been used in their countries to control the presence of those chemicals.

The reported methods have been summarised in the annexed list of "Methods Reported by Member Countries".

In addition three countries commented to the request to provide methods to identify dioxins and related compounds.

Two of these countries expressed their favour in having method criteria which have to be fulfilled by the procedure ("fit-for-purpose") instead of individual accepted methods. Countries referred to the European Community and its Directive 2002/69/EC of 26 July 2002 laying down the sampling methods and the methods of analysis for the official control of dioxins and the determination of dioxin-like PCBs in foodstuffs.

The list in Annex 2 is organised as follows:

Each method is referred to the country reporting it.

For each method a general scope is mentioned if indicated by the notifier (column 2).

The principle of the method used is indicated in column 3.

If available a reference is given in column 4.

Information on the validation status is given in column 5.

**Criteria approach**

The criteria defined (Annex 1) below are based on the validation results of only two methods for the identification of dioxins and related compounds. Due to the lack of sufficient validation data for such methods it is proposed to take into consideration the criteria laid down in Commission Directive 2002/69/EC of 26 July 2002 laying down the sampling methods and the methods of analysis for the official control of dioxins and the determination of dioxin-like PCBs in foodstuffs and Commission Directive 2002/70/EC of 26 July 2002 establishing requirements for the determination of levels of dioxins and dioxin-like PCBs in feeding stuffs (Official Journal of the European Communities L 209, pages 5-14 and 15-21, 6.8.2002), which are already valid in 25 Codex member states and for which some experience exists.

## **ANNEX 1: Method Criteria for Determination of Dioxins/Furans and dioxin-like PCBs**

### **1. Requirements for Confirmatory Methods:**

Note: Confirmatory methods are usually high-resolution gas chromatography/high resolution mass spectrometry methods.

#### 1.1. Applicability:

All foods and feeding stuffs

#### 1.2. Selectivity:

A distinction is required for PCDDs, PCDFs and dioxin-like PCBs from a multitude of other, co-extracted and possibly interfering compounds present at concentrations up to several orders of magnitude higher than those of the analytes of interest. Separation of dioxins from interfering chlorinated compounds such as PCBs and chlorinated diphenyl ethers should be carried out by suitable chromatographic techniques (preferably with a florisil, alumina and/or carbon column). For gas chromatography/mass spectrometry (GC/MS) methods a differentiation among various congeners is necessary, such as between toxic (e.g. the seventeen 2,3,7,8-substituted PCDDs and PCDFs and dioxin-like PCBs) and other congeners.

Gaschromatographic separation of isomers should be sufficient (< 25 % peak to peak between 1,2,3,4,7,8-HxCDF and 1,2,3,6,7,8-HxCDF).

#### 1.3. Limits of detection:

For PCDDs and PCDFs, detectable quantities have to be in the picogram TEQ ( $10^{-12}$  g) range because of extreme toxicity of some of these compounds. PCBs are known to occur at higher levels than the PCDDs and PCDFs. For most PCB congeners sensitivity in the nanogram ( $10^{-9}$  g) range is already sufficient. However, for the measurement of the more toxic dioxin-like PCB congeners (in particular non-ortho substituted congeners), the same sensitivity must be reached as for the PCDDs and PCDFs.

#### 1.4. Limits of quantification, differences between upperbound and lowerbound level:

The accepted specific limit of quantification of an individual congener is the concentration of an analyte in the extract of a sample which produces an instrumental response at two different ions, to be monitored with an S/N (signal/noise) ratio of 3:1 for the less sensitive signal and fulfilment of the basic requirements such as, e.g., retention time, isotope ratio according to the determination procedure as described in EPA method 1613 revision B<sup>9</sup>.

The difference between upperbound level and lower bound level should not exceed 20 % for foodstuffs with a dioxin contamination of about 1 pg WHO-TEQ/g fat (based on PCDD/PCDF only). For foodstuffs with a low fat content, the same requirements for contamination levels of about 1 pg WHO-TEQ/g product have to be applied. For lower contamination levels, for example 0.50 pg WHO-TEQ/g product, the difference between upperbound and lowerbound level may be in the range of 25 to 40 %.

The concept of 'upperbound' requires using the limit of quantification for the contribution of each non-quantified congener to the TEQ.

The concept of 'lowerbound' requires using zero for the contribution of each non-quantified congener to the TEQ.

#### 1.5 Recovery:

Control of recovery is necessary. For confirmatory methods, the recoveries of the individual internal standards should be in the range of 60 % to 120 %. Lower or higher recoveries for individual congeners, in particular for some hepta- and octa- chlorinated dibenzodioxins and dibenzofurans, are acceptable on the

---

<sup>9</sup> Validation data for EPA 1613: Telliard, William A., McCarty, Harry B., and Riddick, Lynn S. "Results of the Interlaboratory Validation Study of US EPA Method 1613 for the Analysis of Tetra- through Octachlorinated Dioxins and Furans by Isotope Dilution GC/MS," *Chemosphere*, 27, 41-46 (1993).

condition that their contribution to the TEQ value does not exceed 10 % of the total TEQ value (based on PCDD/F only).

*Use of internal standards:*

Addition of <sup>13</sup>C-labelled 2,3,7,8-chlorine substituted internal PCDD/F standards (and of <sup>13</sup>C-labelled internal dioxin-like PCB standards, if dioxin-like PCBs have to be determined) must be carried out at the very beginning or start of the analytical method e.g. prior to extraction in order to validate the analytical procedure. At least one congener for each of the tetra- to octa-chlorinated homologous groups for PCDD/F (and at least one congener for each of the homologous groups for dioxin-like PCBs, if dioxin-like PCBs have to be determined) must be added (alternatively, at least one congener for each mass spectrometric selected ion recording function used for monitoring PCDD/F and dioxin-like PCBs). There is a clear preference, certainly in case of confirmatory methods, of using all 17 <sup>13</sup>C-labelled 2,3,7,8-substituted internal PCDD/F standards and all 12 <sup>13</sup>C-labelled internal dioxin-like PCB standard (if dioxin-like PCBs have to be determined). Relative response factors should also be determined for those congeners for which no <sup>13</sup>C-labelled analogue is added by using appropriate calibration solutions.

For foodstuffs of plant origin and foodstuffs of animal origin containing less than 10 % fat, the addition of the internal standards is mandatory prior to extraction. For foodstuffs of animal origin containing more than 10 % fat, the internal standards can be added either before extraction or after fat extraction. The same specifications apply for the analysis of feeding stuff of plant as well as animal origin.

An appropriate validation of the extraction efficiency should be carried out, depending on the stage at which internal standards are introduced and on whether results are reported on product or fat basis. Prior to GC/MS analysis, 1 or 2 recovery (surrogate) standard(s) must be added.

**1.6 Accuracy (trueness and precision):**

The determination should provide a valid estimate of the true concentration in a sample. High accuracy (accuracy of the measurement: the closeness of the agreement between the result of a measurement with the true or assigned value of the measurement) is necessary to avoid the rejection of a sample analysis result on the basis of poor reliability of the estimate of TEQ. Accuracy is expressed as trueness (difference between the mean value measured for an analyte in a certified material and its certified value, expressed as percentage of this value) and precision (RSD<sub>R</sub>, relative standard deviation calculated from results generated under reproducibility conditions). Following criteria have to be complied with on total TEQ value:

	Confirmatory methods
Trueness	- 20 % to + 20 %
Precision RSD <sub>R</sub>	< 15 %

**2. Requirements for Screening-Techniques:**

Note: GC/MS methods of analysis and bioassays may be used for screening. For cell based bioassays specific requirements are laid down in point 2.5 and for kit-based bioassays in point 2.6. Positive results have always to be confirmed by a confirmatory method of analysis (HRGC/HRMS).

**2.1 Applicability:**

All foods and feeding stuffs

**2.2. Selectivity:**

For bioassays, the target compounds, possible interferences and maximum tolerable blank levels should be defined. Bioassays should be able to determine TEQ values selectively as the sum of PCDDs, PCDFs and dioxin-like PCBs.

Information on the number of false-positive and false-negative results of a large set of samples below and above the maximum level or action level is necessary, in comparison to the TEQ content as determined by a confirmatory method of analysis.



The rate of false positive samples should be low enough to make the use of a screening tool advantageous.

A blank and a reference sample(s) have to be included in each test series, which is extracted and tested at the same time under identical conditions. The reference sample must show a clearly elevated response in comparison to a blank.

Extra reference samples  $0.5 \times$  and  $2 \times$  the level of interest should be included to demonstrate the proper performance of the test in the range of interest for the control of the level of interest.

When testing specific matrices, the suitability of the reference sample(s) has to be demonstrated, preferentially by including samples shown by HRGC/HRMS to contain a TEQ level around that of the reference sample or else a blank spiked at this level.

### 2.3 Recovery (GC/MS methods):

For screening methods, the recoveries of the internal standards should be in the range of 30% to 140%. For proper use of internal standards refer to 1.5.

### 2.4. Accuracy in case of a quantitative screening:

The quantitative approach requires standard dilution series, duplicate or triplicate clean up and measuring as well as blank and recovery controls. The result may be expressed as TEQ, thereby assuming that the compounds responsible for the signal correspond to the TEQ principle. This can be performed by using TCDD (or a dioxin/furan standard mixture) to produce a calibration curve to calculate the TEQ level in the extract and thus in the sample. This is subsequently corrected for the TEQ level calculated for a blank sample (to account for impurities from solvents and chemicals used), and a recovery (calculated from the TEQ level in a quality control sample around the level of interest). It is essential to note that part of the apparent recovery loss may be due to matrix effects and/or differences between the TEF values in the bioassays and the official TEF values set by WHO.

Since no internal standards can be used in bioassays, tests on repeatability are very important to obtain information on the standard deviation within one test series. The coefficient of variation should be below 30 %. Actual false negative rates should be under 1%.

	Screening methods
False negative rate	< 1%
Precision $RSD_R$	< 30 %

### 2.5 Specific requirements for cell-based bioassays

- When performing a bioassay, every test run requires a series of reference concentrations of TCDD or a dioxin/furan mixture (full dose-response curve with a  $R^2 > 0.95$ ). However, for screening purposes an expanded low level curve for analysing low level samples could be used.
- A TCDD reference concentration (about  $3 \times$  limit of quantification) on a quality control sheet should be used for the outcome of the bioassay over a constant time period. An alternative could be the relative response of a reference sample in comparison to the TCDD calibration line since the response of the cells may depend on many factors.
- Quality control (QC) charts for each type of reference material should be recorded and checked to make sure the outcome is in accordance with the stated guidelines.
- In particular for quantitative calculations, the induction of the sample dilution used must be within the linear portion of the response curve. Samples above the linear portion of the response curve must be diluted and re-tested. Therefore, at least three or more dilutions at one time are recommended to be tested.
- The percent standard deviation should not be above 15 % in a triplicate determination for each sample dilution and not above 30 % between three independent experiments.

- The limit of detection may be set as  $3\times$  the standard deviation of the solvent blank or of the background response. Another approach is to apply a response that is above the background (induction factor  $5\times$  the solvent blank) calculated from the calibration curve of the day. The limit of quantification may be set as  $5\times$  to  $6\times$  the standard deviation of the solvent blank or of the background response or to apply a response that is above the background (induction factor  $10\times$  the solvent blank) calculated from the calibration curve of the day.
- Information on correspondence between bioassay and HRGC/HRMS results should be made available.

#### 2.6 Specific requirements for kit-based bioassays<sup>10</sup>

- Manufacturer's instructions for sample preparation and analyses have to be followed.
- Test kits should not be used after the expiration date.
- Materials or components designed for use with other kits should not be used.
- Test kits should be kept within the specified range of storage temperature and used at the specified operating temperature.
- The limit of detection for immunoassays is determined as  $3\times$  the standard deviation, based on 10 replicate analysis of the blank, to be divided by the slope value of the linear regression equation.
- Reference standards should be used for tests at the laboratory to make sure that the response to the standard is within an acceptable range.

---

<sup>10</sup> No evidence has yet been submitted of commercially available kit-based bioassays having sufficient sensitivity and reliability to be used for screening for the presence of dioxins at the required levels in samples of foodstuffs and feeding stuffs.

**ANNEX2:Methods reported by governments and organisations**

<b><u>Member state</u></b>	<b><u>Applicability</u></b>	<b><u>Principle</u></b>	<b><u>Reference</u></b>	<b><u>Comment</u></b>
<b>USA</b>	Food	Ion trap	D. G. Hayward, K. Hooper, and D. Andrzejewski. Tandem-in-time mass spectrometry method for the sub-parts-per-trillion determination of 2,3,7,8-chlorine-substituted dibenzo-p-dioxins and -furans in high-fat foods. <i>Analytical Chemistry</i> 71 (1):212-220, 1999.	Not validated
<b>USA</b>	Food	Ion trap HRMS	D. G. Hayward, J. Holcomb, R. Glidden, P. Wilson, M. Harris, and V. Spencer. Quadrupole ion storage tandem mass spectrometry and high-resolution mass spectrometry: complementary application in the measurement of 2,3,7,8-chlorine substituted dibenzo-p-dioxins and dibenzofurans in US foods. <i>Chemosphere</i> 43 (4-7):407-415, 2001.	Not validated
<b>Germany</b>	Feed	HRMS	Determination of PCDDs, PCDFs and selected coplanar(non-ortho-) PCBs in feeding stuffs VDLUFA - Collection of methods,VDLUFA-Verlag Darmstadt, Germany, VDLUFA (1996b) Band VII: Umweltanalytik – Dioxine in Futtermitteln 3.3.2.4.	<b>Validated</b>
<b>Germany</b>	soil, sewage sludge and compost	HRMS	Determination of PCDDs, PCDFs and selected coplanar (non-ortho-) PCBs in soil, sewage sludge and compost VDLUFA - Collection of methods VDLUFA-Verlag Darmstadt, Germany, VDLUFA (1996b) Band VII: Umweltanalytik – Dioxine in Böden, KS und Komposten 3.3.2.3.	<b>Validated</b>
<b>Germany</b>	Food	HRMS	Determination of PCDDs and PCDFs in foods of animal origin, P. Fürst, CVUA Münster, Germany	<b>Validated</b>