

Pathways to precision in soil analysis: advancing soil laboratories in Latin America and the Caribbean

Caminos hacia la
Precisión en el Análisis de
Suelos: avance de los
Laboratorios de Suelos
en América Latina y el
Caribe



LATSOLAN
LATIN AMERICAN SOIL LABORATORY NETWORK

Calibration

Daniel Vidal Pérez

WORKSHOP
SANTIAGO - CHILLÁN | CHILE
8-11 APRIL 2024



- Lab Glassware (volumetric);
- Equipments (Weigh balance, UV/Vis, Flame Photometer, etc);
- Signal x Results – Calibration curve

LAB GLASSWARE (VOLUMETRIC)


- Some examples are pipets, volumetric flasks and burets;
- The value marked on a piece of glassware is called the nominal value;
- The true value is usually slightly different from the nominal value;
- The key measuring device in the laboratory is the analytical balance.

LAB GLASSWARE (VOLUMETRIC)

- To Contain vs. To Deliver;
 - A. The volume of a pipet is determined by weighing the water **delivered** into a clean dry container.
From the weight and the density of water one can calculate the true volume delivered;
 - B. To calibrate a volumetric flask, it is first weighed empty, clean and dry.
It is then **filled** to the mark with water and again weighed.
As before, the volume is calculated from the weight of water and its density

LAB GLASSWARE (VOLUMETRIC)

- An example: Calibration of a Burette




Using the burette, deliver 10.00 mL of the distilled water into the pre-weighed 100-mL volumetric flask.

LAB GLASSWARE (VOLUMETRIC)

- An example: Calibration of a 50 mL Burette

Figure 5. Final Volume: 10.00-mL Reading



Record the mass of the filled volumetric flask.

LAB GLASSWARE (VOLUMETRIC)

- An example: Calibration of a Burette

Density of Water at 25.0 °C (from Table 5): 0.997044 g/mL

Mass of Water and Volumetric Flask (g): 78.4294 g

–

Mass of Empty 100-mL Volumetric Flask (g): – 68.3996 g

Mass of Water (g): = 10.0298 g

$$\text{Density} = \frac{\text{Mass}}{\text{Volume}}$$

$$\text{Volume} = \frac{\text{Mass}}{\text{Density}}$$

$$\text{Volume} = \frac{10.0298\text{g}}{0.997044\text{g/mL}} = 10.0595 \text{ mL}$$

Table 5. Density of Water at Various Temperatures

| Temperature (°C) | Density (g/mL) | Temperature (°C) | Density (g/mL) | Temperature (°C) | Density (g/mL) |
|------------------|----------------|------------------|----------------|------------------|----------------|
| 19.0 | 0.998405 | 21.6 | 0.997860 | 24.2 | 0.997246 |
| 19.2 | 0.998365 | 21.8 | 0.997815 | 24.4 | 0.997196 |
| 19.4 | 0.998325 | 22.0 | 0.997770 | 24.6 | 0.997146 |
| 19.6 | 0.998285 | 22.2 | 0.997724 | 24.8 | 0.997095 |
| 19.8 | 0.998265 | 22.4 | 0.997678 | 25.0 | 0.997044 |
| 20.0 | 0.998203 | 22.6 | 0.997632 | 25.2 | 0.996992 |
| 20.2 | 0.998162 | 22.8 | 0.997585 | 25.4 | 0.996941 |
| 20.4 | 0.998120 | 23.0 | 0.997538 | 25.6 | 0.996888 |
| 20.6 | 0.998078 | 23.2 | 0.997490 | 25.8 | 0.996836 |
| 20.8 | 0.998035 | 23.4 | 0.997442 | 26.0 | 0.996783 |
| 21.0 | 0.997922 | 23.6 | 0.997394 | 26.2 | 0.996729 |
| 21.2 | 0.997948 | 23.8 | 0.997345 | 26.4 | 0.996703 |
| 21.4 | 0.997904 | 24.0 | 0.997296 | | |

EQUIPMENTS (WEIGH BALANCE)

- Factors that affect the performance of analytical balances:
 1. acceleration of gravity;
 2. ambient temperature;
 3. humidity;
 4. barometric pressure;
 5. altitude, etc

- Conclusion: Calibration can only be given at the place of installation

EQUIPMENTS (WEIGH ANALYTICAL BALANCE)

- External calibration is a manual process that requires the user to place an **approved set of weighing scale weights** on the balance and set their weight as the standard;
- **Performance Check:** 1 mg, 2 mg, 5 mg, 10 mg and 20 mg calibration standard weights should be placed separately on the top pan balance after the auto calibration process, and the measurements noted in the performance check log.
- To meet the traceable standard, these measurements need to remain within the 0.1% of the actual mass value of each weight.



Pathways to precision in soil analysis: advancing soil laboratories in Latin America and the Caribbean

Caminos hacia la Precisión en el Análisis de Suelos: avance de los Laboratorios de Suelos en América Latina y el Caribe

WORKSHOP | SANTIAGO - CHILLÁN | CHILE


8-11 APRIL 2024



Signal x Results – CALIBRATION CURVES

- UV-Vis spectrophotometer example (Absorbance reading)

| Abs | Abs Corr | Conc |
|---------|----------|------|
| 0,03591 | 0 | 0 |
| 0,168 | 0,13209 | 10 |
| 0,3068 | 0,27089 | 25 |
| 0,5009 | 0,46499 | 50 |
| 0,7341 | 0,69819 | 70 |



Signal x Results – CALIBRATION CURVES

- Calibration curve MUST include ZERO?

| RESUMO DOS RESULTADOS | | | | | | |
|---------------------------------|----------------------|--------------------|---------------|----------------|--------------------------|-----------------------|
| <i>estatística de regressão</i> | | | | | | |
| R múltiplo | 0,996815 | | | | | |
| R-Quadrado | 0,99364 | | | | | |
| R-quadrado | 0,991519 | | | | | |
| Erro padrão | 2,653078 | | | | | |
| Observação | 5 | | | | | |
| ANOVA | | | | | | |
| | <i>gl</i> | <i>SQ</i> | <i>MQ</i> | <i>F</i> | <i>F de significação</i> | |
| Regressão | 1 | 3298,883532 | 3298,884 | 468,6698 | 0,000215697 | |
| Resíduo | 3 | 21,11646806 | 7,038823 | | | |
| Total | 4 | 3320 | | | | |
| | <i>Coefficientes</i> | <i>Erro padrão</i> | <i>Stat t</i> | <i>valor-P</i> | <i>95% inferiores</i> | <i>95% superiores</i> |
| Interseção | -1,63276 | 1,918315169 | -0,85114 | 0,457236 | -7,737696562 | 4,472184937 |
| Abs Corr | 104,1808 | 4,812316125 | 21,64878 | 0,000216 | 88,86583349 | 119,4957376 |

Conc = 104.18 Abs – 1.63
Can I force to pass 0.00?

Signal x Results – CALIBRATION CURVES

- Calibration curve could include ZERO

| RESUMO DOS RESULTADOS | | | | | | |
|---------------------------------|---------------------|--------------------|---------------|----------------|--------------------------|-----------------------|
| <i>estatística de regressão</i> | | | | | | |
| R múltiplo | 0,996044 | | | | | |
| R-Quadrado | 0,992104 | | | | | |
| R-quadrado | 0,742104 | | | | | |
| Erro padrão | 2,560062 | | | | | |
| Observação | 5 | | | | | |
| ANOVA | | | | | | |
| | <i>gl</i> | <i>SQ</i> | <i>MQ</i> | <i>F</i> | <i>F de significação</i> | |
| Regressão | 1 | 3293,784326 | 3293,784 | 502,5672 | 0,000194347 | |
| Resíduo | 4 | 26,21567433 | 6,553919 | | | |
| Total | 5 | 3320 | | | | |
| | <i>Coefficiente</i> | <i>Erro padrão</i> | <i>Stat t</i> | <i>valor-P</i> | <i>95% inferiores</i> | <i>95% superiores</i> |
| Interseção | 0 | #N/D | #N/D | #N/D | #N/D | #N/D |
| Abs Corr | 100,9623 | 2,872100989 | 35,15276 | 3,91E-06 | 92,9880218 | 108,9365163 |

Figures and Illustrations References

- <https://web.pdx.edu/~atkinsdb/teach/321/CALVOL.pdf>
- https://www.kbcc.cuny.edu/academicdepartments/physci/documents/chemistry/Chem11_Lab2_CalibrationGlassware_V4_03-04-2022.pdf
- <https://www.depts.ttu.edu/meatscience/docs/WhyCalibrateAnalyticalBalancesandScales-6-30-20.pdf>

